



Declaration of Conformity

The manufacturer, A&D Company, Limited, is exclusively responsible for the declaration of conformity.

Manufacturer 's Name and Address:

A&D Company, Limited
1-243 Asahi, Kitamoto-shi, Saitama-ken. 364-8585 Japan

European Representative:

A&D INSTRUMENTS LIMITED
Unit24/26 Blacklands Way
Abingdon Business Park
Abingdon, Oxon
OX14 1DY United Kingdom

Product Category:

Blood pressure monitor

Model Number:

UA-651BLE

Classification/Rule:

II a/ Rule10

- ✓ **The product complies with the requirements of the Directive 93/42/EEC Annex I and the certified quality system according to Annex II excluding (4).**

Standards:

EN 60601-1:2006/A1:2013, EN 60601-1-2:2015, EN 60601-1-6:2010,
EN 60601-1-11: 2010, EN 62304:2006, EN 62366:2008,
IEC 80601-2-30:2009/A1:2013, ISO 81060-2:2013, EN ISO 10993-1:2009,
EN ISO 14971:2012, EN 1041:2008, EN ISO 15223-1:2016

Notified Body:

TÜV SÜD Product Service GmbH (0123)
Ridlerstraße 65 80339 MÜNCHEN, GERMANY

Certificate number: G1 085349 0008 Rev. 00

This declaration is valid until the expiration date of the MDD annex II certificate by TÜV SÜD Product Service GmbH.

✓ **The product complies with the requirements of the Directive 2014/53/EU.**

Standards:

EN 300 328 V2.1.1


Notified Body: Not Applied

✓ **The product complies with the requirements of the Directive 2011/65/EU.**

Notified Body: Not Applied

Place and date: Saitama, Japan, 14 Mar. 2019

File number: 19031410



Minoru Saitou

Chief of Quality Control Office

A&D Company, Limited